



**SVENSK FÖRENING FÖR RADIOFYSIK**  
Swedish Society of Radiation Physics  
(Member of IUPESM)

# **Image quality**

## **Technical/physical aspects**


Nationella kvalitetsdokument för digital radiologi  
AG1

Michael Sandborg och Jalil Bahar  
Radiofysikavdelningen Linköping

# Requirements on QA-tests

---

- Objective
- High precision
- Quick and simple
- Available
- (Universal and automatic)




# Inspiration from



---

Statens strålskyddsinstitut  
Projekt P1069.98

## Utvärdering av datortomografers doseffektivitet

Peter Hägglund, Rolf Johansson och Göran Wickman  
Institutionen för strålningsvetenskaper Umeå Universitet






# Method

---


- Objective measures to characterise imaging systems (or detector) performance exists
  - i.e. MTF, NPS, NEQ and DQE
  - Difficult to adapt and time-consuming
- but simple visual evaluation of test phantoms are often used in practice ...
  - # of visible low- and high-contrast details
  - Simple but rather imprecise (subjective)



# Suggestion

---

- **LDI: Low-contrast Detection Index**  
is
  - Objective
  - Precise
  - Reproducible
  - Semi-quick?




# Material

---


- Leeds To10 phantom (SN174)
- Philips Digital Diagnost (2001)
- 4 x 5 cm PMMA blocks
- Barracuda MPD kerma-meter
- Same image processing (Unique) 'Bäcken'
- FDA=100 cm
- 25 x 25 cm field size
- No table top cushion
- Manuel exposure

# Material

- Leeds To10 phantom (SN174)
- Philips Digital Diagnost (2001)
- 4 x 5 cm PMMA blocks
- Barracuda MPD kerma-meter
- Same image processing (Unique) 'Bäcken'
- FDA=100 cm
- 25 x 25 cm field size
- No table top cushion
- Manuel exposure







SON

NO.2

NO.1

SON



POLONIUS, OFELIA,  
450406-9248

C: 19343.5, W: 2447.0  
S= 800.0

↔ 13.75mm

2007-03-12, 15:48:03  
US Linköping, lab16

BÄCKEN FRONTAL



Zoom to pixel-to-pixel



POLONIUS, OFELIA,  
450406-9248

C: 19343.5, W: 2447.0  
S= 800.0

Medel: 21506, Std avv: 328



2007-03-12, 15:48:03  
US Linköping, lab16

BÄCKEN FRONTAL



A circular ROI is applied. Average p.v. and s.d. is measured

POLONIUS, OFELIA,  
450406-9248

C: 20063.5, W: 2447.0  
S= 400.0

Medel: 19791, Std.avv.: 263

Medel: 20875, Std.avv.: 282

Medel: 19707, Std.avv.: 261

Medel: 19444, Std.avv.: 262

BÄCKEN FRONTAL



2007-03-12, 14:30:14  
US Linköping, lab16



Measure p.v. and s.d. on top and beside the detail



# Method

Uncorrected



# Method




Uncorrected






# Method

Uncorrected linear reg, LDI



# Method



Corrected linear reg, LDI







# Method

Corrected linear reg, LDI



# Method

Corrected log reg, LDI



# Repeat 3-5 times

POLONIUS, OFELIA,  
450406-9248

Session 2, 15 bitar, 60kV, 32mAs, 2mmAl, 1,23mGy, Bild#1  
C: 20723.5, W: 6847.0  
S= 800.0

POLONIUS, OFELIA,  
450406-9248

Session 2, 15 bitar, 60kV, 32mAs, 2mmAl, 1,23mGy, Bild#2  
C: 20523.5, W: 4847.0  
S= 800.0

POLONIUS, OFELIA,  
450406-9248

Session 2, 15 bitar, 60kV, 32mAs, 2mmAl, 1,23mGy, Bild#3  
C: 20723.5, W: 5407.0  
S= 800.0

POLONIUS, OFELIA,  
450406-9248

Session 2, 15 bitar, 60kV, 32mAs, 2mmAl, 1,23mGy, Bild#4  
C: 20923.5, W: 5247.0  
S= 800.0

POLONIUS, OFELIA,  
450406-9248

Session 2, 15 bitar, 60kV, 32mAs, 2mmAl, 1,23mGy, Bild#5  
C: 20323.5, W: 3407.0  
S= 800.0


2007-03-12, 15  
US Linköping, I

2007-03-12, 15  
US Linköping, I

2007-03-12, 15  
US Linköping, I

2007-03-12, 15  
US Linköping, I



2007-03-12, 15:48:03  
US Linköping, lab16





BÄCKEN FRONTAL

# Method


5 image samples



# Results: Uncertainty




# Results: Uncertainty



# Uncertainty


---



# Uncertainty

---

- Uncertainty is stable with
  - Bit-depth
  - Detector dose






# Uncertainty

---


- Uncertainty is stable with
  - Bit-depth
  - Detector dose
- **BUT very important** to set up and align the phantom and x-ray unit in a consistent way!



# Uncertainty


---

- Uncertainty is stable with
  - Bit-depth
  - Detector dose
- **BUT very important** to set up and align the phantom and x-ray unit in a consistent way!
- **Keep detailed notations** and document x-ray unit set up (photograph)



# Uncertainty


- Uncertainty is stable with
  - Bit-depth
  - Detector dose
- **BUT very important** to set up and align the phantom and x-ray unit in a consistent way!
- **Keep detailed notations** and document x-ray unit set up (photograph)





# Applications / tests of method

---



- LDIs dependence on
  - bit-depth
  - kV
  - ESAK





# Results: Different Bit depth



# Results: Different kV





# Results: Different ESAK



# Results: Correlations

Subjective v.s. Objective assessment






# Conclusions and further work

---


- LDI method seem ´appropriate´ for QA-test
- Further tests and validations of sensitivity to different
  - image processing schemes
  - dynamic range of image data
  - X-ray units
  - sizes of contrast detail
  - test phantoms
- Automatic image assessment with software



# Conclusions and further work


---

- LDI method seem ´appropriate´ for QA-test
- Further tests and validations of sensitivity to different
  - image processing schemes
  - dynamic range of image data
  - X-ray units
  - sizes of contrast detail
  - test phantoms
- Automatic image assessment with software



# Comments ? !

---



# Comments ? !

---

- Please send comments and suggestions for improvement to
  - [Jalil.Bahar@lio.se](mailto:Jalil.Bahar@lio.se) or
  - [Michael.Sandborg@lio.se](mailto:Michael.Sandborg@lio.se)

